

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
 Data File : VN085942.D
 Acq On : 11 Mar 2025 19:29
 Operator : JC\MD
 Sample : PB167050ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167050ZHE#09

Quant Time: Mar 12 01:28:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	145638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	294524	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	273505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	98482	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	122146	64.691	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 129.380%#	
35) Dibromofluoromethane	8.165	113	103786	53.853	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.700%	
50) Toluene-d8	10.565	98	375171	53.505	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.020%	
62) 4-Bromofluorobenzene	12.847	95	107397	46.487	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 92.980%	
Target Compounds						
						Qvalue
16) Acetone	4.424	43	28527	47.748	ug/l	96
20) Methylene Chloride	5.277	84	5344	2.781	ug/l	93
43) Isopropyl Acetate	8.682	43	79463	19.843	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
Data File : VN085942.D
Acq On : 11 Mar 2025 19:29
Operator : JC\MD
Sample : PB167050ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167050ZHE#09

Quant Time: Mar 12 01:28:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

