

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
 Data File : VN085946.D
 Acq On : 11 Mar 2025 21:05
 Operator : JC\MD
 Sample : Q1534-24
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-636-COMP-19

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/12/2025
 Supervised By : Mahesh Dadoda 03/12/2025

Quant Time: Mar 12 01:30:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	154562	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302708	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	291370	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	110784	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	119680	59.725	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 119.460%			
35) Dibromofluoromethane	8.165	113	100523	50.749	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.500%			
50) Toluene-d8	10.565	98	371260	51.516	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.040%			
62) 4-Bromofluorobenzene	12.847	95	115045	48.452	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.900%			
Target Compounds					Qvalue	
16) Acetone	4.424	43	41349	65.214	ug/l	93
20) Methylene Chloride	5.277	84	6575	3.225	ug/l	95
43) Isopropyl Acetate	8.688	43	46625m	11.091	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
Data File : VN085946.D
Acq On : 11 Mar 2025 21:05
Operator : JC\MD
Sample : Q1534-24
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-636-COMP-19

Quant Time: Mar 12 01:30:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 03/12/2025
Supervised By :Mahesh Dadoda 03/12/2025

