

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
 Data File : VN085947.D
 Acq On : 11 Mar 2025 21:29
 Operator : JC\MD
 Sample : Q1514-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ENV-105-SB01

Quant Time: Mar 12 01:31:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	154053	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	299260	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	282311	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	99543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118390	59.277	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 118.560%		
35) Dibromofluoromethane	8.165	113	99800	50.965	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.920%		
50) Toluene-d8	10.565	98	368578	51.733	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.460%		
62) 4-Bromofluorobenzene	12.847	95	111523	47.509	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 95.020%		
Target Compounds						
						Qvalue
16) Acetone	4.424	43	25696	40.661	ug/l	95
20) Methylene Chloride	5.271	84	5536	2.724	ug/l #	88
43) Isopropyl Acetate	8.683	43	71882	17.605	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
Data File : VN085947.D
Acq On : 11 Mar 2025 21:29
Operator : JC\MD
Sample : Q1514-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ENV-105-SB01

Quant Time: Mar 12 01:31:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

