

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
 Data File : VN085948.D
 Acq On : 11 Mar 2025 21:53
 Operator : JC\MD
 Sample : Q1514-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ENV-105-SB02

Quant Time: Mar 12 01:31:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	150437	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	296435	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	287045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	110049	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116697	59.833	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	119.660%
35) Dibromofluoromethane	8.159	113	98753	50.911	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.820%
50) Toluene-d8	10.565	98	358054	50.735	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.480%
62) 4-Bromofluorobenzene	12.847	95	114862	49.398	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.800%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	23134	37.486	ug/l	99
20) Methylene Chloride	5.283	84	5982	3.014	ug/l #	85
43) Isopropyl Acetate	8.688	43	70964	17.544	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031125\
 Data File : VN085948.D
 Acq On : 11 Mar 2025 21:53
 Operator : JC\MD
 Sample : Q1514-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ENV-105-SB02

Quant Time: Mar 12 01:31:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

