

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
 Data File : VN054172.D
 Acq On : 12 Mar 2019 19:08
 Operator : JC/SP
 Sample : K1837-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MR-MH-07-COMP

Manual Integrations
 APPROVED

MMDadoda
 3/13/2019 11:42:20 AM

Quant Time: Mar 13 08:03:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1489414	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2422109	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2503705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	1081090	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	793468	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
35) Dibromofluoromethane	7.59	113	750456	48.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.08%	
50) Toluene-d8	10.09	98	3027407	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
62) 4-Bromofluorobenzene	12.40	95	1149189	56.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.86%	
Target Compounds						
16) Acetone	3.82	43	28180	6.222	ug/l	90
20) Methylene Chloride	4.55	84	22688	1.526	ug/l	95
25) 2-Butanone	6.88	43	9957m	1.873	ug/l	
43) Isopropyl Acetate	8.17	43	50143	2.229	ug/l	97
95) Naphthalene	15.13	128	9236	3.651	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
Data File : VN054172.D
Acq On : 12 Mar 2019 19:08
Operator : JC/SP
Sample : K1837-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MR-MH-07-COMP

Manual Integrations
APPROVED

MMDadoda
3/13/2019 11:42:20 AM

Quant Time: Mar 13 08:03:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

