

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
 Data File : VN054174.D
 Acq On : 12 Mar 2019 20:01
 Operator : JC/SP
 Sample : K1837-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MR-MH-11-COMP

Quant Time: Mar 13 08:07:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1444732	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2396685	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2468827	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	1101828	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	789492	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
35) Dibromofluoromethane	7.59	113	738802	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
50) Toluene-d8	10.09	98	2953412	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
62) 4-Bromofluorobenzene	12.40	95	1140660	57.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.20%	
Target Compounds						
16) Acetone	3.82	43	30839	7.020	ug/l	92
20) Methylene Chloride	4.54	84	21042	1.459	ug/l	92
43) Isopropyl Acetate	8.17	43	48946	2.199	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
Data File : VN054174.D
Acq On : 12 Mar 2019 20:01
Operator : JC/SP
Sample : K1837-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MR-MH-11-COMP

Quant Time: Mar 13 08:07:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

