

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
 Data File : VN054182.D
 Acq On : 13 Mar 2019 00:24
 Operator : JC/SP
 Sample : K1849-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 H1308-PC

Quant Time: Mar 13 08:25:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1361228	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2269464	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2418736	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	1194182	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	761779	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
35) Dibromofluoromethane	7.59	113	702845	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
50) Toluene-d8	10.09	98	2858518	52.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.76%	
62) 4-Bromofluorobenzene	12.40	95	1157888	61.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.44%	
Target Compounds						
16) Acetone	3.82	43	61757	14.920	ug/l	95
20) Methylene Chloride	4.54	84	1222480	89.947	ug/l	99
43) Isopropyl Acetate	8.17	43	48678	2.309	ug/l	98
95) Naphthalene	15.13	128	282936	8.497	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
Data File : VN054182.D
Acq On : 13 Mar 2019 00:24
Operator : JC/SP
Sample : K1849-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
H1308-PC

Quant Time: Mar 13 08:25:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

