

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
 Data File : VN054186.D
 Acq On : 13 Mar 2019 2:09
 Operator : JC/SP
 Sample : K1849-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PUMP-CON

Quant Time: Mar 13 08:34:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1442035	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2455407	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2512515	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	1097975	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	807446	52.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.44%	
35) Dibromofluoromethane	7.59	113	761519	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
50) Toluene-d8	10.09	98	3032863	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
62) 4-Bromofluorobenzene	12.40	95	1140308	55.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.44%	
Target Compounds						
16) Acetone	3.82	43	31068	7.085	ug/l	93
20) Methylene Chloride	4.54	84	62123	4.315	ug/l	93
43) Isopropyl Acetate	8.17	43	47909	2.101	ug/l	98
95) Naphthalene	15.13	128	25115	3.955	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
Data File : VN054186.D
Acq On : 13 Mar 2019 2:09
Operator : JC/SP
Sample : K1849-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PUMP-CON

Quant Time: Mar 13 08:34:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

