

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
 Data File : VN054197.D
 Acq On : 13 Mar 2019 6:59
 Operator : JC/SP
 Sample : K1837-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MR-MH-10-COMP

Quant Time: Mar 13 09:04:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1366248	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2369244	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2561746	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	1195027	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	785001	54.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.20%	
35) Dibromofluoromethane	7.59	113	731723	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
50) Toluene-d8	10.09	98	2972068	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
62) 4-Bromofluorobenzene	12.40	95	1205340	61.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.08%	
Target Compounds						
16) Acetone	3.82	43	31553	7.595	ug/l	94
20) Methylene Chloride	4.55	84	22612	1.658	ug/l	87
43) Isopropyl Acetate	8.17	43	48311	2.195	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
Data File : VN054197.D
Acq On : 13 Mar 2019 6:59
Operator : JC/SP
Sample : K1837-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MR-MH-10-COMP

Quant Time: Mar 13 09:04:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

