

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
 Data File : VN054198.D
 Acq On : 13 Mar 2019 7:26
 Operator : JC/SP
 Sample : K1849-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C1355-PC

Quant Time: Mar 13 09:09:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1426061	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2346887	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2511592	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	1203917	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	776514	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
35) Dibromofluoromethane	7.59	113	723150	47.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.56%	
50) Toluene-d8	10.09	98	2915755	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
62) 4-Bromofluorobenzene	12.40	95	1187143	60.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.38%	
Target Compounds						
16) Acetone	3.82	43	102162	23.560	ug/l	97
20) Methylene Chloride	4.55	84	47350	3.326	ug/l	96
25) 2-Butanone	6.86	43	14745	2.897	ug/l	94
43) Isopropyl Acetate	8.17	43	46955	2.154	ug/l	98
95) Naphthalene	15.13	128	64264	4.602	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
Data File : VN054198.D
Acq On : 13 Mar 2019 7:26
Operator : JC/SP
Sample : K1849-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C1355-PC

Quant Time: Mar 13 09:09:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

