

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
 Data File : VN054200.D
 Acq On : 13 Mar 2019 8:19
 Operator : JC/SP
 Sample : K1687-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HR-06-031119-C

Quant Time: Mar 13 09:15:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1360768	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2260143	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2414616	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	1181712	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	744704	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
35) Dibromofluoromethane	7.59	113	696125	47.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.52%	
50) Toluene-d8	10.09	98	2816853	52.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.64%	
62) 4-Bromofluorobenzene	12.40	95	1149309	61.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.02%	
Target Compounds						
16) Acetone	3.82	43	36247	8.760	ug/l	95
20) Methylene Chloride	4.55	84	1028179	75.676	ug/l	99
43) Isopropyl Acetate	8.17	43	47764	2.275	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	89413	1.159	ug/l	99
95) Naphthalene	15.13	128	49657	4.361	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031219\
Data File : VN054200.D
Acq On : 13 Mar 2019 8:19
Operator : JC/SP
Sample : K1687-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HR-06-031119-C

Quant Time: Mar 13 09:15:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

