

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031220\
 Data File : VN060478.D
 Acq On : 12 Mar 2020 14:48
 Operator : JC/MD
 Sample : PB127467ZHE#15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127467ZHE#15

Quant Time: Mar 13 03:52:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	191545	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	320505	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	304410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142227	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	126800	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
35) Dibromofluoromethane	7.57	113	95456	64.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.32%	
50) Toluene-d8	10.08	98	396114	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
62) 4-Bromofluorobenzene	12.40	95	149157	52.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.52%	
Target Compounds						
16) Acetone	3.80	43	6213	7.829	ug/l	97
20) Methylene Chloride	4.52	84	3302	1.515	ug/l #	83
43) Isopropyl Acetate	8.14	43	66510	13.135	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031220\
Data File : VN060478.D
Acq On : 12 Mar 2020 14:48
Operator : JC/MD
Sample : PB127467ZHE#15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127467ZHE#15

Quant Time: Mar 13 03:52:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

