

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031220\
 Data File : VN060479.D
 Acq On : 12 Mar 2020 15:11
 Operator : JC/MD
 Sample : PB127467ZHE#16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127467ZHE#16

Quant Time: Mar 13 03:53:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	191930	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	320183	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	302489	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	138595	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	125285	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
35) Dibromofluoromethane	7.57	113	94754	63.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.52%	
50) Toluene-d8	10.08	98	393324	52.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.58%	
62) 4-Bromofluorobenzene	12.40	95	146753	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
Target Compounds						
16) Acetone	3.79	43	6491	8.163	ug/l	94
20) Methylene Chloride	4.52	84	3021	1.383	ug/l #	91
43) Isopropyl Acetate	8.14	43	65703	12.989	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031220\
Data File : VN060479.D
Acq On : 12 Mar 2020 15:11
Operator : JC/MD
Sample : PB127467ZHE#16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127467ZHE#16

Quant Time: Mar 13 03:53:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

