

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031220\
 Data File : VN060483.D
 Acq On : 12 Mar 2020 16:41
 Operator : JC/MD
 Sample : PB127465ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127465ZHE#05

Quant Time: Mar 13 04:00:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	189954	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	316773	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	299423	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	138389	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	124125	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
35) Dibromofluoromethane	7.57	113	94125	64.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.04%	
50) Toluene-d8	10.08	98	391723	52.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.28%	
62) 4-Bromofluorobenzene	12.40	95	148091	52.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.00%	
Target Compounds						
16) Acetone	3.79	43	6838	8.688	ug/l	100
20) Methylene Chloride	4.53	84	3911	1.809	ug/l	95
43) Isopropyl Acetate	8.14	43	58060	11.602	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031220\
Data File : VN060483.D
Acq On : 12 Mar 2020 16:41
Operator : JC/MD
Sample : PB127465ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127465ZHE#05

Quant Time: Mar 13 04:00:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

