

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031220\
 Data File : VN060485.D
 Acq On : 12 Mar 2020 17:27
 Operator : JC/MD
 Sample : PB127465ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127465ZHE#07

Quant Time: Mar 13 04:04:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	190369	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	313484	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	296975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	138629	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	125522	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
35) Dibromofluoromethane	7.56	113	94275	64.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	129.58%	
50) Toluene-d8	10.08	98	388662	52.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.56%	
62) 4-Bromofluorobenzene	12.40	95	146196	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
Target Compounds						
16) Acetone	3.79	43	6744	8.550	ug/l	100
20) Methylene Chloride	4.53	84	4525	2.089	ug/l	96
43) Isopropyl Acetate	8.14	43	59483	12.011	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031220\
Data File : VN060485.D
Acq On : 12 Mar 2020 17:27
Operator : JC/MD
Sample : PB127465ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127465ZHE#07

Quant Time: Mar 13 04:04:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

