

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031221\
 Data File : VN066184.D
 Acq On : 12 Mar 2021 14:54
 Operator : JC/MD
 Sample : VN0312WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0312WBL01

Quant Time: Mar 13 02:40:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 12 13:00:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	438072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	684735	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	559118	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	200564	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	377538	53.41	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.82%
35) Dibromofluoromethane	7.507	113	259744	56.22	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	112.44%
50) Toluene-d8	10.028	98	967875	51.73	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.46%
62) 4-Bromofluorobenzene	12.345	95	285923	41.54	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	83.08%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031221\
Data File : VN066184.D
Acq On : 12 Mar 2021 14:54
Operator : JC/MD
Sample : VN0312WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0312WBL01

Quant Time: Mar 13 02:40:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031221W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 12 13:00:23 2021
Response via : Initial Calibration

