

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031224\
 Data File : VN081369.D
 Acq On : 12 Mar 2024 11:21
 Operator : JC\MD
 Sample : VN0312WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0312WBL01

Quant Time: Mar 13 04:49:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	332196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	596612	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	508676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	198832	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	244001	50.800	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.600%
35) Dibromofluoromethane	8.165	113	190143	52.019	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.040%
50) Toluene-d8	10.571	98	713928	52.248	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.500%
62) 4-Bromofluorobenzene	12.853	95	226383	46.876	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	93.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031224\
Data File : VN081369.D
Acq On : 12 Mar 2024 11:21
Operator : JC\MD
Sample : VN0312WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0312WBL01

Quant Time: Mar 13 04:49:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

