

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031224\
 Data File : VN081375.D
 Acq On : 12 Mar 2024 16:14
 Operator : JC\MD
 Sample : P1734-03 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1340

Quant Time: Mar 13 04:53:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	339618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	597980	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	519590	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	213357	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	245387	49.972	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.940%
35) Dibromofluoromethane	8.165	113	188012	51.319	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.640%
50) Toluene-d8	10.570	98	705313	51.500	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.000%
62) 4-Bromofluorobenzene	12.853	95	229005	47.310	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	94.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031224\
Data File : VN081375.D
Acq On : 12 Mar 2024 16:14
Operator : JC\MD
Sample : P1734-03 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1340

Quant Time: Mar 13 04:53:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

