

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031225\
 Data File : VN085955.D
 Acq On : 12 Mar 2025 13:05
 Operator : JC\MD
 Sample : VN0312MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0312MBL01

Quant Time: Mar 13 01:33:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	154850	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	280779	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245080	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	100106	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	122536	61.037	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	122.080%
35) Dibromofluoromethane	8.165	113	102616	55.852	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.700%
50) Toluene-d8	10.565	98	339112	50.730	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.460%
62) 4-Bromofluorobenzene	12.847	95	103833	47.145	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031225\
Data File : VN085955.D
Acq On : 12 Mar 2025 13:05
Operator : JC\MD
Sample : VN0312MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0312MBL01

Quant Time: Mar 13 01:33:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

