

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031225\
 Data File : VN085961.D
 Acq On : 12 Mar 2025 15:42
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 13 01:37:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	136653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	254348	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	223256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	91011	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	111944	63.186	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	126.380%#
35) Dibromofluoromethane	8.165	113	93162	55.976	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.960%
50) Toluene-d8	10.565	98	308992	51.028	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.060%
62) 4-Bromofluorobenzene	12.847	95	95242	47.738	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.480%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031225\
Data File : VN085961.D
Acq On : 12 Mar 2025 15:42
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 13 01:37:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085961.D
Acq: 12 Mar 2025 15:42




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 136653
Ion Ratio Lower Upper
168 100
99 62.7 47.9 71.9



#33
1,2-Dichloroethane-d4
Concen: 63.186 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.006 min
Lab File: VN085961.D
Acq: 12 Mar 2025 15:42

Tgt Ion: 65 Resp: 111944
Ion Ratio Lower Upper
65 100
67 51.9 0.0 106.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN085961.D

Acq: 12 Mar 2025 15:42

Instrument :

MSVOA_N

ClientSampleId :

IBLK





Tgt Ion:114 Resp: 254348

Ion Ratio Lower Upper

114 100

63 21.8 0.0 39.2

88 16.7 0.0 30.0



#35

Dibromofluoromethane

Concen: 55.976 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085961.D

Acq: 12 Mar 2025 15:42




Tgt Ion:113 Resp: 93162


Ion Ratio Lower Upper

113 100

111 103.8 82.2 123.4

192 19.3 14.7 22.1









#50
Toluene-d8
Concen: 51.028 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.006 min
Lab File: VN085961.D
Acq: 12 Mar 2025 15:42




Instrument :
MSVOA_N
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 308992
Ion Ratio Lower Upper
98 100
100 64.5 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 47.738 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.006 min
Lab File: VN085961.D
Acq: 12 Mar 2025 15:42

Tgt Ion: 95 Resp: 95242
Ion Ratio Lower Upper
95 100
174 78.3 0.0 152.4
176 73.8 0.0 146.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085961.D
Acq: 12 Mar 2025 15:42

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Tgt Ion:117 Resp: 223256
Ion Ratio Lower Upper
117 100
82 57.5 45.7 68.5
119 32.1 26.2 39.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.006 min
Lab File: VN085961.D
Acq: 12 Mar 2025 15:42

Tgt Ion:152 Resp: 91011
Ion Ratio Lower Upper
152 100
115 61.6 30.4 91.3
150 159.0 0.0 345.4

