

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031319\
 Data File : VN054242.D
 Acq On : 14 Mar 2019 5:17
 Operator : JC/SP
 Sample : K1866-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Mar 14 08:55:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1338438	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2346336	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2428528	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	1083708	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	768824	54.09	ug/l	0.00
Spiked Amount			Recovery	=	108.18%	
35) Dibromofluoromethane	7.59	113	725420	47.94	ug/l	0.00
Spiked Amount			Recovery	=	95.88%	
50) Toluene-d8	10.09	98	2891271	51.73	ug/l	0.00
Spiked Amount			Recovery	=	103.46%	
62) 4-Bromofluorobenzene	12.40	95	1108173	56.67	ug/l	0.00
Spiked Amount			Recovery	=	113.34%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	3.41	74	2241	0.263	ug/l #	22
10) Methyl Iodide	3.95	142	163	1.505	ug/l #	45
11) Tert butyl alcohol	4.81	59	210	0.216	ug/l #	70
13) Acrolein	3.61	56	222	0.240	ug/l #	14
16) Acetone	3.82	43	30659	7.533	ug/l	90
20) Methylene Chloride	4.55	84	37651	2.817	ug/l	97
25) 2-Butanone	6.86	43	4099	0.858	ug/l	97
29) Tetrahydrofuran	7.23	42	2436	0.793	ug/l #	80
31) Cyclohexane	7.66	56	20713	Below Cal	#	3
36) 1,1-Dichloropropene	7.66	75	81468	3.770	ug/l #	48
38) Carbon Tetrachloride	7.66	117	115479	5.096	ug/l #	17
43) Isopropyl Acetate	8.17	43	43149	1.980	ug/l	95
48) Methyl methacrylate	9.18	41	18629	1.684	ug/l #	11
49) 1,4-Dioxane	9.20	88	63	0.403	ug/l #	1
51) 4-Methyl-2-Pentanone	10.03	43	18912	1.640	ug/l #	36
53) t-1,3-Dichloropropene	10.38	75	235	1.995	ug/l #	42
54) cis-1,3-Dichloropropene	9.90	75	173253	7.149	ug/l #	61
57) 1,3-Dichloropropane	10.76	76	31005	1.325	ug/l #	43
59) 2-Hexanone	10.77	43	398171	51.886	ug/l #	50
71) Bromoform	12.40	173	6967	2.800	ug/l #	1
74) N-amyl acetate	11.88	43	39479	2.051	ug/l #	53
76) 1,2,3-Trichloropropane	12.40	75	531812	33.371	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	531812	114.077	ug/l #	11
93) 1,2,4-Trichlorobenzene	14.90	180	564	2.540	ug/l #	40
95) Naphthalene	15.14	128	2338	3.515	ug/l #	76


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031319\
Data File : VN054242.D
Acq On : 14 Mar 2019 5:17
Operator : JC/SP
Sample : K1866-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Mar 14 08:55:32 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17

Instrument :

MSVOA_N


ClientSampled :

Tot Ion:168 Resp: 1338438

Ion Ratio Lower Upper

168 100

99 48.7 37.1 55.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

600000

500000

400000



300000

200000

100000

0

Time-->



#8

Diethyl Ether

Concen: 0.263 ug/l

RT: 3.41 min Scan# 509

Delta R.T. -0.00 min

Lab File: VN054242.D


Acq: 14 Mar 2019 5:17

Tgt Ion: 74 Resp: 2241

Ion Ratio Lower Upper

74 100

45 14.4 43.1 129.4#



Abundance Ion 74.00 (73.70 to 74.70): VNO

Ion 45.00 (44.70 to 45.70): VNO


3000


2000

1000

0

Time-->





#10

Methvl Iodide

Concen: 1.505 ug/l

RT: 3.95 min Scan# 676

Delta R.T. 0.00 min


Lab File: VN054242.D

Acq: 14 Mar 2019 5:17

Instrument :
MSVOA_N
ClientSampled :

Tot Ion:142 Resp: 163




Ion	Ratio	Lower	Upper
142	100		
127	0.0	30.3	45.5#
141	0.0	11.2	16.8#



Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V



#11

Tert butyl alcohol

Concen: 0.216 ug/l

RT: 4.81 min Scan# 945


Delta R.T. 0.00 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17



Tgt Ion: 59 Resp: 210


Ion	Ratio	Lower	Upper
59	100		
57	0.0	9.0	13.6#



Abundance Ion 59.00 (58.70 to 59.70): V

Ion 57.00 (56.70 to 57.70): V





#13

Acrolein

Concen: 0.240 ug/l

RT: 3.61 min Scan# 571

Delta R.T. 0.00 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17

Instrument :

MSVOA_N





ClientSampled :

Tot Ion: 56 Resp: 222

Ion Ratio Lower Upper

56 100

55 0.0 56.7 85.1#



#16

Acetone

Concen: 7.533 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.00 min

Lab File: VN054242.D




Acq: 14 Mar 2019 5:17


Tgt Ion: 43 Resp: 30659

Ion Ratio Lower Upper

43 100

58 43.5 29.9 44.9









#20
Methvlene Chloride
Concen: 2.817 ug/l
RT: 4.55 min Scan# 862
Delta R.T. 0.00 min
Lab File: VN054242.D
Acq: 14 Mar 2019 5:17




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 84 Resp: 37651
Ion Ratio Lower Upper
84 100
49 110.5 89.7 134.5
51 36.4 28.2 42.4
86 68.8 50.8 76.2



#25
2-Butanone
Concen: 0.858 ug/l
RT: 6.86 min Scan# 1581
Delta R.T. 0.01 min
Lab File: VN054242.D
Acq: 14 Mar 2019 5:17

Tgt Ion: 43 Resp: 4099
Ion Ratio Lower Upper
43 100
72 27.4 23.0 34.6









#29
Tetrahydrofuran
Concen: 0.793 ug/l
RT: 7.23 min Scan# 1698
Delta R.T. 0.01 min
Lab File: VN054242.D
Acq: 14 Mar 2019 5:17




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 42 Resp: 2436
Ion Ratio Lower Upper
42 100
72 29.4 37.8 56.6#
71 53.7 36.2 54.2



#31
Cyclohexane
Concen: Below Cal
RT: 7.66 min Scan# 1832
Delta R.T. 0.01 min
Lab File: VN054242.D
Acq: 14 Mar 2019 5:17

Tgt Ion: 56 Resp: 20713
Ion Ratio Lower Upper
56 100
69 170.5 27.8 41.6#
84 136.7 74.6 111.8#





#33

1,2-Dichloroethane-d4

Concen: 54.088 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion: 65 Resp: 768824

Ion Ratio Lower Upper




65 100

67 52.0 0.0 104.6



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. 0.00 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17


Tgt Ion: 114 Resp: 2346336

Ion Ratio Lower Upper

114 100

63 17.8 0.0 35.2



88 15.0 0.0 30.2




Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN0

Ion 87.90 (87.60 to 88.60): VN0









#35
 Dibromofluoromethane
 Concen: 47.937 ug/l
 RT: 7.59 min Scan# 1809
 Delta R.T. 0.00 min
 Lab File: VN054242.D
 Acq: 14 Mar 2019 5:17




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
113	100		
111	102.7	82.2	123.4
192	23.7	19.2	28.8



#36
 1,1-Dichloropropene
 Concen: 3.770 ug/l
 RT: 7.66 min Scan# 1832
 Delta R.T. -0.13 min
 Lab File: VN054242.D
 Acq: 14 Mar 2019 5:17

Tgt Ion	Ratio	Lower	Upper
75	100		
110	8.6	19.1	57.5#
77	0.0	25.0	37.4#





#38

Carbon Tetrachloride

Concen: 5.096 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.11 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17

Instrument :

MSVOA_N

ClientSampled :


Tot Ion:117 Resp: 115479

Ion Ratio Lower Upper

117 100

119 5.7 76.3 114.5#




121 0.0 24.6 36.8#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 1.980 ug/l

RT: 8.17 min Scan# 1990

Delta R.T. 0.00 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17


Tgt Ion: 43 Resp: 43149

Ion Ratio Lower Upper

43 100

61 23.0 15.8 23.8



87 15.6 12.0 18.0




Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 61.00 (60.70 to 61.70): VNO

Ion 87.00 (86.70 to 87.70): VNO





#48

Methyl methacrylate

Concen: 1.684 ug/l

RT: 9.18 min Scan# 2304

Delta R.T. -0.02 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion: 41 Resp: 18629

Ion Ratio Lower Upper




41 100

69 0.6 72.6 108.8#

39 0.0 47.5 71.3#



Abundance Ion 41.00 (40.70 to 41.70): VNO
Ion 69.00 (68.70 to 69.70): VNO
Ion 39.00 (38.70 to 39.70): VNO



#49

1,4-Dioxane

Concen: 0.403 ug/l

RT: 9.20 min Scan# 2311

Delta R.T. 0.00 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17


Tgt Ion: 88 Resp: 63

Ion Ratio Lower Upper



88 100


43 0.0 24.1 36.1#

58 3157.1 50.8 76.2#



Abundance Ion 88.00 (87.70 to 88.70): VNO
Ion 43.00 (42.70 to 43.70): VNO
Ion 58.00 (57.70 to 58.70): VNO





#50

Toluene-d8

Concen: 51.733 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion: 98 Resp: 2891271

Ion Ratio Lower Upper




98 100

100 63.5 51.2 76.8



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN



#51

4-Methyl-2-Pentanone

Concen: 1.640 ug/l

RT: 10.03 min Scan# 2567

Delta R.T. 0.04 min

Lab File: VN054242.D


Acq: 14 Mar 2019 5:17

Tgt Ion: 43 Resp: 18912

Ion Ratio Lower Upper



43 100


58 0.0 31.8 47.6#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0





#53

t-1,3-Dichloropropene

Concen: 1.995 ug/l

RT: 10.38 min Scan# 2678

Delta R.T. 0.00 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion: 75 Resp: 235

Ion Ratio Lower Upper




75 100

77 0.0 25.8 38.8#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 76.90 (76.60 to 77.60): VN0



#54

cis-1,3-Dichloropropene

Concen: 7.149 ug/l

RT: 9.90 min Scan# 2529

Delta R.T. 0.06 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17


Tgt Ion: 75 Resp: 173253

Ion Ratio Lower Upper

75 100

77 0.5 25.4 38.0#



39 28.2 37.2 55.8#




Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 76.90 (76.60 to 77.60): VN0

Ion 39.00 (38.70 to 39.70): VN0









#57
 1,3-Dichloropropane
 Concen: 1.325 ug/l
 RT: 10.76 min Scan# 2796
 Delta R.T. 0.06 min
 Lab File: VN054242.D
 Acq: 14 Mar 2019 5:17




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 76 Resp: 31005
 Ion Ratio Lower Upper
 76 100
 78 0.1 25.7 38.5#



#59
 2-Hexanone
 Concen: 51.886 ug/l
 RT: 10.77 min Scan# 2797
 Delta R.T. 0.01 min
 Lab File: VN054242.D
 Acq: 14 Mar 2019 5:17

Tgt Ion: 43 Resp: 398171
 Ion Ratio Lower Upper
 43 100
 58 19.5 28.0 83.9#









#62
4-Bromofluorobenzene
Concen: 56.669 ug/l
RT: 12.40 min Scan# 3306
Delta R.T. 0.00 min
Lab File: VN054242.D
Acq: 14 Mar 2019 5:17




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 95 Resp: 1108173
Ion Ratio Lower Upper
95 100
174 97.2 0.0 191.0
176 93.9 0.0 187.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN054242.D
Acq: 14 Mar 2019 5:17

Tgt Ion: 117 Resp: 2428528
Ion Ratio Lower Upper
117 100
82 50.8 40.0 60.0
119 32.5 25.8 38.6





#71

Bromoform

Concen: 2.800 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.28 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:173 Resp: 6967

Ion Ratio Lower Upper




173 100

175 932.2 24.4 73.2#

254 0.0 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17


Tgt Ion:152 Resp: 1083708

Ion Ratio Lower Upper



152 100


115 54.9 27.2 81.6

150 157.7 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#74

N-amvl acetate

Concen: 2.051 ug/l

RT: 11.88 min Scan# 3143

Delta R.T. -0.19 min

Lab File: VN054242.D


Acq: 14 Mar 2019 5:17

Instrument :

MSVOA_N

ClientSampleId :

Tot Ion:	43	Resp:	39479
Ion	Ratio	Lower	Upper
43	100		
70	2.6	35.4	53.0#
55	6.1	19.8	29.6#
61	10.9	19.8	29.6#






Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 70.00 (69.70 to 70.70): VN0

Ion 55.00 (54.70 to 55.70): VN0

Ion 61.00 (60.70 to 61.70): VN0



#76

1,2,3-Trichloropropane

Concen: 33.371 ug/l


RT: 12.40 min Scan# 3306

Delta R.T. -0.15 min

Lab File: VN054242.D



Acq: 14 Mar 2019 5:17


Tgt Ion:	75	Resp:	531812
Ion	Ratio	Lower	Upper
75	100		
77	1.2	0.0	0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0









#81
trans-1,4-Dichloro-2-butene
Concen: 114.077 ug/l
RT: 12.40 min Scan# 3306
Delta R.T. 0.10 min
Lab File: VN054242.D
Acq: 14 Mar 2019 5:17




Instrument :
MSVOA_N
ClientSampleId :


Tot Ion: 75 Resp: 531812
Ion Ratio Lower Upper
75 100
53 0.0 76.2 114.2#
89 0.0 39.2 58.8#



#93
1,2,4-Trichlorobenzene
Concen: 2.540 ug/l
RT: 14.90 min Scan# 4084
Delta R.T. -0.00 min
Lab File: VN054242.D
Acq: 14 Mar 2019 5:17

Tgt Ion: 180 Resp: 564
Ion Ratio Lower Upper
180 100
182 35.8 48.0 144.2#
145 0.0 14.2 42.8#





#95

Naphthalene

Concen: 3.515 ug/l

RT: 15.14 min Scan# 4156

Delta R.T. 0.00 min

Lab File: VN054242.D

Acq: 14 Mar 2019 5:17

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:128 Resp: 2338

Ion Ratio Lower Upper

128 100

127 4.5 10.1 15.1#

129 0.0 8.7 13.1#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

