

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031319\
 Data File : VN054214.D
 Acq On : 13 Mar 2019 16:07
 Operator : JC/SP
 Sample : K1842-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-MW-10_030719

Quant Time: Mar 14 07:23:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1388195	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2298319	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2371178	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	990773	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	745598	50.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.14%	
35) Dibromofluoromethane	7.59	113	704302	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
50) Toluene-d8	10.09	98	2841363	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
62) 4-Bromofluorobenzene	12.40	95	1057351	55.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.40%	
Target Compounds						
30) Chloroform	7.37	83	29891	1.208	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031319\
Data File : VN054214.D
Acq On : 13 Mar 2019 16:07
Operator : JC/SP
Sample : K1842-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-MW-10_030719

Quant Time: Mar 14 07:23:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

