

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031319\
 Data File : VN054243.D
 Acq On : 14 Mar 2019 5:43
 Operator : JC/SP
 Sample : K1866-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MR-MH-14-COMP

Quant Time: Mar 14 09:02:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1388197	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2309682	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2452656	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	1131624	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	752320	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
35) Dibromofluoromethane	7.59	113	708457	47.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.12%	
50) Toluene-d8	10.09	98	2889591	52.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.04%	
62) 4-Bromofluorobenzene	12.40	95	1124573	58.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.84%	
Target Compounds						
16) Acetone	3.83	43	34962	8.283	ug/l	95
20) Methylene Chloride	4.55	84	38226	2.758	ug/l	96
43) Isopropyl Acetate	8.17	43	42382	1.976	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031319\
Data File : VN054243.D
Acq On : 14 Mar 2019 5:43
Operator : JC/SP
Sample : K1866-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MR-MH-14-COMP

Quant Time: Mar 14 09:02:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

