

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031319\
 Data File : VN054248.D
 Acq On : 14 Mar 2019 7:55
 Operator : JC/SP
 Sample : K1866-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MR-MH-04-COMP

Quant Time: Mar 14 09:11:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1432216	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2422870	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2580315	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	1180130	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	785447	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
35) Dibromofluoromethane	7.59	113	738834	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
50) Toluene-d8	10.09	98	3020592	52.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.68%	
62) 4-Bromofluorobenzene	12.40	95	1191766	59.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.04%	
Target Compounds						
16) Acetone	3.82	43	32545	7.473	ug/l	93
20) Methylene Chloride	4.54	84	39987	2.796	ug/l	95
43) Isopropyl Acetate	8.17	43	44110	1.960	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031319\
Data File : VN054248.D
Acq On : 14 Mar 2019 7:55
Operator : JC/SP
Sample : K1866-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MR-MH-04-COMP

Quant Time: Mar 14 09:11:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

