

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031320\
 Data File : VN060507.D
 Acq On : 13 Mar 2020 16:57
 Operator : JC/MD
 Sample : L1761-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-01-031220


Quant Time: Mar 16 04:30:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	203473	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	341364	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	324139	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	153668	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.00	65	133370	51.09	ug/l	0.00
Spiked Amount			Recovery	=	102.18%	
35) Dibromofluoromethane	7.57	113	100899	63.68	ug/l	0.00
Spiked Amount			Recovery	=	127.36%	
50) Toluene-d8	10.08	98	422898	52.73	ug/l	0.00
Spiked Amount			Recovery	=	105.46%	
62) 4-Bromofluorobenzene	12.40	95	161555	53.14	ug/l	0.00
Spiked Amount			Recovery	=	106.28%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.79	43	8789	10.425	ug/l	97
18) Methyl Acetate	4.29	43	1221	0.434	ug/l #	52
20) Methylene Chloride	4.52	84	5225	2.257	ug/l	88
25) 2-Butanone	6.81	43	1352	0.993	ug/l #	49
29) Tetrahydrofuran	7.18	42	218	0.230	ug/l #	33
31) Cyclohexane	7.64	56	4547	Below Cal	#	1
36) 1,1-Dichloropropene	7.64	75	15410	4.737	ug/l #	50
37) Ethyl Acetate	6.81	43	1352	0.439	ug/l #	68
38) Carbon Tetrachloride	7.64	117	19884	6.277	ug/l #	16
43) Isopropyl Acetate	8.14	43	61254	11.358	ug/l #	91
45) 1,2-Dichloropropane	9.25	63	1104	0.446	ug/l #	44
48) Methyl methacrylate	9.16	41	3202	1.302	ug/l #	13
51) 4-Methyl-2-Pentanone	9.97	43	5109	1.671	ug/l	96
54) cis-1,3-Dichloropropene	9.89	75	26181	6.311	ug/l #	59
57) 1,3-Dichloropropane	10.75	76	3748	0.961	ug/l #	42
59) 2-Hexanone	10.75	43	58532	26.071	ug/l #	52
68) m/p-Xylenes	11.61	106	1324	0.294	ug/l	89
69) o-Xylene	11.94	106	862	0.201	ug/l	69
74) N-amyl acetate	12.02	43	3474	0.698	ug/l #	44
76) 1,2,3-Trichloropropane	12.40	75	81601	24.203	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	3866	0.403	ug/l	90
81) trans-1,4-Dichloro-2-buten	12.40	75	81601	66.812	ug/l #	8
83) tert-Butylbenzene	13.03	119	1729	0.208	ug/l #	73
84) 1,2,4-Trimethylbenzene	13.04	105	13050	1.348	ug/l	97
90) Hexachloroethane	13.89	117	360	0.242	ug/l #	8
95) Naphthalene	15.13	128	9006	0.716	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1826

Delta R.T. 0.00 min

Lab File: VN060507.D

Acq: 13 Mar 2020 16:57

Instrument :

MSVOA_N

ClientSampleId :


SU-01-031220

Tot Ion:168 Resp: 203473

Ion Ratio Lower Upper




168 100

99 57.2 44.7 67.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN



#16

Acetone

Concen: 10.425 ug/l

RT: 3.79 min Scan# 628

Delta R.T. 0.01 min

Lab File: VN060507.D


Acq: 13 Mar 2020 16:57

Tgt Ion: 43 Resp: 8789

Ion Ratio Lower Upper



43 100


58 29.3 24.6 36.8



Abundance Ion 43.00 (42.70 to 43.70): VN

Ion 58.00 (57.70 to 58.70): VN









#18
Methyl Acetate
Concen: 0.434 ug/l
RT: 4.29 min Scan# 784
Delta R.T. 0.01 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57




Instrument :
MSVOA_N
ClientSampled :
SU-01-031220


Tot Ion: 43 Resp: 1221
Ion Ratio Lower Upper
43 100
74 0.0 18.8 28.2#



#20
Methylene Chloride
Concen: 2.257 ug/l
RT: 4.52 min Scan# 855
Delta R.T. 0.00 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57

Tgt Ion: 84 Resp: 5225
Ion Ratio Lower Upper
84 100
49 108.0 99.3 148.9
51 35.5 30.1 45.1
86 54.0 51.1 76.7









#25
2-Butanone
Concen: 0.993 ug/l
RT: 6.81 min Scan# 1566
Delta R.T. 0.01 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57




Instrument :
MSVOA_N
ClientSampleId :
SU-01-031220


Tot Ion: 43 Resp: 1352
Ion Ratio Lower Upper
43 100
72 0.0 20.9 31.3#



#29
Tetrahydrofuran
Concen: 0.230 ug/l
RT: 7.18 min Scan# 1683
Delta R.T. 0.01 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57

Tgt Ion: 42 Resp: 218
Ion Ratio Lower Upper
42 100
72 0.0 34.6 52.0#
71 0.0 32.9 49.3#





#33

1,2-Dichloroethane-d4

Concen: 51.088 ug/l

RT: 8.00 min Scan# 1938

Delta R.T. 0.00 min

Lab File: VN060507.D

Acq: 13 Mar 2020 16:57

Instrument :

MSVOA_N

ClientSampleId :


SU-01-031220

Tot Ion: 65 Resp: 133370

Ion Ratio Lower Upper

65 100

67 52.7 0.0 103.8



Abundance Ion 64.90 (64.60 to 65.60): VN060507.D (-1845) (-)

Ion 66.90 (66.60 to 67.60): VN060507.D (-1845) (-)

8.00

60000

50000

40000

30000



20000

10000

0

7.90 8.00 8.10

Time-->



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.57 min Scan# 2113

Delta R.T. 0.00 min

Lab File: VN060507.D

Acq: 13 Mar 2020 16:57


Tgt Ion: 114 Resp: 341364

Ion Ratio Lower Upper

114 100

63 22.3 0.0 43.2

88 16.6 0.0 31.6



Abundance Ion 113.90 (113.60 to 114.60): VN060507.D (-2020) (-)

Ion 63.00 (62.70 to 63.70): VN060507.D (-2020) (-)

Ion 87.90 (87.60 to 88.60): VN060507.D (-2020) (-)

8.57

200000

150000


100000


50000

0

8.50 8.60 8.70

Time-->





#35

Dibromofluoromethane

Concen: 63.679 ug/l

RT: 7.57 min Scan# 1802

Delta R.T. 0.00 min

Lab File: VN060507.D

Acq: 13 Mar 2020 16:57

Instrument :

MSVOA_N

ClientSampleId :

SU-01-031220





Tot Ion:113 Resp: 100899

Ion Ratio Lower Upper

113 100

111 104.2 82.4 123.6

192 20.1 16.2 24.2



#36

1,1-Dichloropropene

Concen: 4.737 ug/l

RT: 7.64 min Scan# 1825

Delta R.T. -0.13 min

Lab File: VN060507.D

Acq: 13 Mar 2020 16:57




Tgt Ion: 75 Resp: 15410


Ion Ratio Lower Upper

75 100

110 9.0 17.3 52.0#

77 0.0 24.2 36.2#









#37
Ethyl Acetate
Concen: 0.439 ug/l
RT: 6.81 min Scan# 1566
Delta R.T. -0.09 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57

Instrument :
MSVOA_N
ClientSampleId :
SU-01-031220

Tot Ion: 43 Resp: 1352
Ion Ratio Lower Upper
43 100
61 0.0 11.1 16.7#
70 0.0 8.4 12.6#




Abundance Ion 43.00 (42.70 to 43.70): VN0
Ion 60.90 (60.60 to 61.60): VN0
Ion 70.00 (69.70 to 70.70): VN0






#38
Carbon Tetrachloride
Concen: 6.277 ug/l
RT: 7.64 min Scan# 1826
Delta R.T. -0.11 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57

Tgt Ion: 117 Resp: 19884
Ion Ratio Lower Upper
117 100
119 4.6 76.7 115.1#
121 0.0 24.5 36.7#



Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V





#43

Isopropyl Acetate

Concen: 11.358 ug/l

RT: 8.14 min Scan# 1980

Delta R.T. 0.00 min

Lab File: VN060507.D

Acq: 13 Mar 2020 16:57

Instrument :


MSVOA_N

ClientSampleId :




SU-01-031220

Tot Ion: 43 Resp: 61254

Ion	Ratio	Lower	Upper
43	100		
61	20.1	20.9	31.3#
87	12.1	10.6	16.0



Abundance Ion 43.00 (42.70 to 43.70): VN060264.D (-1959) (-)



#45

1,2-Dichloropropane

Concen: 0.446 ug/l

RT: 9.25 min Scan# 2325


Delta R.T. 0.15 min

Lab File: VN060507.D



Acq: 13 Mar 2020 16:57


Tgt Ion: 63 Resp: 1104

Ion	Ratio	Lower	Upper
63	100		
65	0.0	24.4	36.6#



Abundance Ion 62.90 (62.60 to 63.60): VN060264.D (-2258) (-)









#48
Methvl methacrylate
Concen: 1.302 ug/l
RT: 9.16 min Scan# 2298
Delta R.T. -0.02 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57

Instrument :
MSVOA_N
ClientSampleId :
SU-01-031220

Tot Ion: 41 Resp: 3202
Ion Ratio Lower Upper
41 100
69 0.0 66.7 100.1#
39 0.0 50.1 75.1#




Abundance Ion 41.00 (40.70 to 41.70): VNO
Ion 69.00 (68.70 to 69.70): VNO
Ion 39.00 (38.70 to 39.70): VNO






#50
Toluene-d8
Concen: 52.735 ug/l
RT: 10.08 min Scan# 2583
Delta R.T. 0.00 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57

Tgt Ion: 98 Resp: 422898
Ion Ratio Lower Upper
98 100
100 65.0 52.2 78.2



Abundance Ion 98.00 (97.70 to 98.70): VNO
Ion 100.00 (99.70 to 100.70): VNO









#51
 4-Methyl-2-Pentanone
 Concen: 1.671 ug/l
 RT: 9.97 min Scan# 2549
 Delta R.T. 0.00 min
 Lab File: VN060507.D
 Acq: 13 Mar 2020 16:57




Instrument :
 MSVOA_N
 ClientSampleId :
 SU-01-031220


Tot Ion: 43 Resp: 5109
 Ion Ratio Lower Upper
 43 100
 58 36.1 30.9 46.3



#54
 cis-1,3-Dichloropropene
 Concen: 6.311 ug/l
 RT: 9.89 min Scan# 2524
 Delta R.T. 0.06 min
 Lab File: VN060507.D
 Acq: 13 Mar 2020 16:57

Tgt Ion: 75 Resp: 26181
 Ion Ratio Lower Upper
 75 100
 77 0.0 24.8 37.2#
 39 32.0 43.8 65.8#





#57

1,3-Dichloropropane

Concen: 0.961 ug/l

RT: 10.75 min Scan# 2793

Delta R.T. 0.06 min

Lab File: VN060507.D

Acq: 13 Mar 2020 16:57

Instrument :

MSVOA_N

ClientSampled :


SU-01-031220

Tot Ion: 76 Resp: 3748




Ion Ratio Lower Upper

76 100

78 0.0 26.2 39.2#



Abundance Ion 75.90 (75.60 to 76.60): VN060507.D (-2682) (-)



#59

2-Hexanone

Concen: 26.071 ug/l

RT: 10.75 min Scan# 2793

Delta R.T. 0.01 min

Lab File: VN060507.D


Acq: 13 Mar 2020 16:57

Tgt Ion: 43 Resp: 58532



Ion Ratio Lower Upper


43 100

58 18.8 26.5 79.5#



Abundance Ion 43.00 (42.70 to 43.70): VN060507.D (-2696) (-)









#62
4-Bromofluorobenzene
Concen: 53.144 ug/l
RT: 12.40 min Scan# 3304
Delta R.T. 0.00 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57

Instrument :
MSVOA_N
ClientSampled :
SU-01-031220

Tot Ion	Ratio	Lower	Upper
95	100		
174	79.0	0.0	159.2
176	76.1	0.0	153.8




Abundance Ion 94.90 (94.60 to 95.60): VN060507.D (-3290) (-)
Ion 173.80 (173.50 to 174.50): VN060507.D (-3290) (-)
Ion 175.80 (175.50 to 176.50): VN060507.D (-3290) (-)






#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.40 min Scan# 2994
Delta R.T. 0.00 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57

Tgt Ion	Ratio	Lower	Upper
117	100		
82	57.8	45.8	68.6
119	31.7	25.8	38.8



Abundance Ion 116.90 (116.60 to 117.60): VN060507.D (-2979) (-)
Ion 82.00 (81.70 to 82.70): VN060507.D (-2979) (-)
Ion 118.90 (118.60 to 119.60): VN060507.D (-2979) (-)







#68
m/p-Xylenes
Concen: 0.294 ug/l
RT: 11.61 min Scan# 3060
Delta R.T. -0.00 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57

Instrument :
MSVOA_N
ClientSampleId :
SU-01-031220



Tot Ion:106 Resp: 1324
Ion Ratio Lower Upper
106 100
91 222.3 164.5 246.7



Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VN




Time-->




#69
o-Xylene
Concen: 0.201 ug/l
RT: 11.94 min Scan# 3163
Delta R.T. 0.00 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57


Tgt Ion:106 Resp: 862
Ion Ratio Lower Upper
106 100
91 269.1 109.5 328.4




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VN



Time-->





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3597

Delta R.T. 0.00 min

Lab File: VN060507.D

Acq: 13 Mar 2020 16:57

Instrument :

MSVOA_N

ClientSampleId :

SU-01-031220


Tot Ion:152 Resp: 153668

Ion Ratio Lower Upper

152 100

115 59.7 30.0 90.0




150 156.3 0.0 349.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 0.698 ug/l

RT: 12.02 min Scan# 3188

Delta R.T. -0.04 min

Lab File: VN060507.D

Acq: 13 Mar 2020 16:57

Tgt Ion: 43 Resp: 3474


Ion Ratio Lower Upper

43 100

70 0.0 32.2 48.2#

55 0.0 18.2 27.4#

61 0.0 19.2 28.8#






Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V









#76
 1,2,3-Trichloropropane
 Concen: 24.203 ug/l
 RT: 12.40 min Scan# 3304
 Delta R.T. -0.15 min
 Lab File: VN060507.D
 Acq: 13 Mar 2020 16:57




Instrument :
 MSVOA_N
 ClientSampleId :
 SU-01-031220


Tot Ion: 75 Resp: 81601
 Ion Ratio Lower Upper
 75 100
 77 1.2 0.0 0.0#



#80
 1,3,5-Trimethylbenzene
 Concen: 0.403 ug/l
 RT: 12.73 min Scan# 3407
 Delta R.T. 0.00 min
 Lab File: VN060507.D
 Acq: 13 Mar 2020 16:57

Tgt Ion: 105 Resp: 3866
 Ion Ratio Lower Upper
 105 100
 120 55.4 24.1 72.4









#81
trans-1,4-Dichloro-2-butene
Concen: 66.812 ug/l
RT: 12.40 min Scan# 3304
Delta R.T. 0.10 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57




Instrument :
MSVOA_N
ClientSampleId :
SU-01-031220


Tot Ion: 75 Resp: 81601
Ion Ratio Lower Upper
75 100
53 0.0 73.8 110.6#
89 0.0 57.1 85.7#



#83
tert-Butylbenzene
Concen: 0.208 ug/l
RT: 13.03 min Scan# 3502
Delta R.T. 0.04 min
Lab File: VN060507.D
Acq: 13 Mar 2020 16:57

Tgt Ion: 119 Resp: 1729
Ion Ratio Lower Upper
119 100
91 80.6 33.1 99.2
134 0.0 12.4 37.0#





#84

1,2,4-Trimethylbenzene

Concen: 1.348 ug/l

RT: 13.04 min Scan# 3503

Delta R.T. 0.00 min

Lab File: VN060507.D

Acq: 13 Mar 2020 16:57

Instrument :

MSVOA_N

ClientSampleId :


SU-01-031220

Tot Ion:105 Resp: 13050

Ion Ratio Lower Upper




105 100

120 42.5 22.3 66.9



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#90

Hexachloroethane

Concen: 0.242 ug/l

RT: 13.89 min Scan# 3767

Delta R.T. 0.01 min

Lab File: VN060507.D


Acq: 13 Mar 2020 16:57

Tgt Ion:117 Resp: 360

Ion Ratio Lower Upper



117 100


201 0.0 41.0 123.0#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V





#95

Naphthalene

Concen: 0.716 ug/l

RT: 15.13 min Scan# 4154

Delta R.T. 0.00 min

Lab File: VN060507.D

Acq: 13 Mar 2020 16:57

Instrument :

MSVOA_N

ClientSampleId :

SU-01-031220



Tot Ion:128 Resp: 9006

Ion Ratio Lower Upper

128 100

127 11.7 10.4 15.6

129 9.7 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

