

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031320\
 Data File : VN060494.D
 Acq On : 13 Mar 2020 12:03
 Operator : JC/MD
 Sample : L1762-13ME
 Misc : 4.33g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-031220ME

Quant Time: Mar 16 09:38:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	187155	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	307617	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	293487	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	145127	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	119079	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
35) Dibromofluoromethane	7.56	113	88293	61.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.68%	
50) Toluene-d8	10.08	98	380914	52.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.42%	
62) 4-Bromofluorobenzene	12.40	95	149012	54.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.80%	
Target Compounds						
52) Toluene	10.14	92	24058	4.488	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031320\
Data File : VN060494.D
Acq On : 13 Mar 2020 12:03
Operator : JC/MD
Sample : L1762-13ME
Misc : 4.33g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-031220ME

Quant Time: Mar 16 09:38:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

