

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031320\
 Data File : VN060497.D
 Acq On : 13 Mar 2020 13:11
 Operator : JC/MD
 Sample : PB127465ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127465ZHE#10

Quant Time: Mar 16 08:42:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	191211	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	315680	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	297162	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	138907	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	124296	50.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.34%	
35) Dibromofluoromethane	7.57	113	93030	63.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	126.98%	
50) Toluene-d8	10.08	98	387423	52.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.48%	
62) 4-Bromofluorobenzene	12.40	95	144844	51.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.04%	
Target Compounds						
16) Acetone	3.79	43	5920	7.473	ug/l	91
20) Methylene Chloride	4.53	84	4282	1.968	ug/l	96
43) Isopropyl Acetate	8.14	43	67591	13.553	ug/l #	90
51) 4-Methyl-2-Pentanone	9.97	43	5460	1.931	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031320\
Data File : VN060497.D
Acq On : 13 Mar 2020 13:11
Operator : JC/MD
Sample : PB127465ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127465ZHE#10

Quant Time: Mar 16 08:42:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

