

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031320\
 Data File : VN060503.D
 Acq On : 13 Mar 2020 15:27
 Operator : JC/MD
 Sample : PB127506ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127506ZHE#04

Quant Time: Mar 16 08:59:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	202514	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	336753	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	324004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	152054	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	133102	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
35) Dibromofluoromethane	7.57	113	100820	64.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	129.00%	
50) Toluene-d8	10.08	98	416222	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
62) 4-Bromofluorobenzene	12.40	95	160487	53.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.04%	
Target Compounds						
16) Acetone	3.80	43	6384	7.608	ug/l	98
20) Methylene Chloride	4.52	84	3954	1.716	ug/l	91
43) Isopropyl Acetate	8.14	43	67013	12.596	ug/l #	90
51) 4-Methyl-2-Pentanone	9.97	43	6175	2.047	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031320\
Data File : VN060503.D
Acq On : 13 Mar 2020 15:27
Operator : JC/MD
Sample : PB127506ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127506ZHE#04

Quant Time: Mar 16 08:59:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

