

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN031320\
 Data File : VN060508.D
 Acq On : 13 Mar 2020 17:20
 Operator : JC/MD
 Sample : L1762-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-031220

Manual Integrations
 APPROVED

MMDadoda
 3/16/2020 2:04:53 PM

Quant Time: Mar 16 09:15:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	205248	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	342193	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	321626	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	150673	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	135241	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
35) Dibromofluoromethane	7.56	113	101506	63.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.82%	
50) Toluene-d8	10.08	98	423319	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
62) 4-Bromofluorobenzene	12.40	95	159612	52.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.76%	
Target Compounds						
16) Acetone	3.79	43	8272	9.727	ug/l	99
20) Methylene Chloride	4.53	84	5348m	2.290	ug/l	
43) Isopropyl Acetate	8.14	43	65343	12.087	ug/l #	90
51) 4-Methyl-2-Pentanone	9.97	43	5501	1.795	ug/l	90
52) Toluene	10.14	92	25478	4.273	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN031320\
Data File : VN060508.D
Acq On : 13 Mar 2020 17:20
Operator : JC/MD
Sample : L1762-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-031220

Manual Integrations
APPROVED

MMDadoda
3/16/2020 2:04:53 PM

Quant Time: Mar 16 09:15:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

