

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076952.D
 Acq On : 13 Mar 2023 11:08
 Operator : JC\MD
 Sample : VN0313WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0313WBL01

Quant Time: Mar 14 06:58:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	536670	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	893115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	766033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	318273	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	353599	44.152	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.300%
35) Dibromofluoromethane	8.172	113	275819	46.338	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.680%
50) Toluene-d8	10.572	98	1038584	47.344	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.680%
62) 4-Bromofluorobenzene	12.854	95	367832	45.842	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076952.D
Acq On : 13 Mar 2023 11:08
Operator : JC\MD
Sample : VN0313WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0313WBL01

Quant Time: Mar 14 06:58:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

