

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076956.D
 Acq On : 13 Mar 2023 12:55
 Operator : JC\MD
 Sample : 01842-02
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-228

Quant Time: Mar 13 13:12:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.225	168	387502	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.101	114	630769	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	586498	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	253442	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.578	65	255916	44.256	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.520%
35) Dibromofluoromethane	8.172	113	190635	45.348	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.700%
50) Toluene-d8	10.566	98	764332	49.333	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.660%
62) 4-Bromofluorobenzene	12.854	95	280725	49.537	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076956.D
Acq On : 13 Mar 2023 12:55
Operator : JC\MD
Sample : 01842-02
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-228

Quant Time: Mar 13 13:12:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

