

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076962.D
 Acq On : 13 Mar 2023 15:20
 Operator : JC\MD
 Sample : 01880-02 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE115D2-20230309

Quant Time: Mar 14 07:03:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	436788	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	713778	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	638457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	263265	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	295735	45.371	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.740%
35) Dibromofluoromethane	8.172	113	224476	47.188	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.380%
50) Toluene-d8	10.572	98	870666	49.661	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.320%
62) 4-Bromofluorobenzene	12.854	95	309232	48.222	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.440%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.390	101	25339	4.837	ug/l	89
12) 1,1-Dichloroethene	4.354	96	13795	2.857	ug/l	97
20) Methylene Chloride	5.301	84	2087	0.359	ug/l #	85
24) 1,1-Dichloroethane	6.584	63	3176	0.324	ug/l #	84
44) Trichloroethene	9.360	130	567614	101.275	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076962.D
Acq On : 13 Mar 2023 15:20
Operator : JC\MD
Sample : 01880-02 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE115D2-20230309

Quant Time: Mar 14 07:03:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

