

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076964.D
 Acq On : 13 Mar 2023 16:08
 Operator : JC\MD
 Sample : 01880-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE115D1-20230309

Quant Time: Mar 14 07:03:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	462940	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	742805	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	652894	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	265270	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	313980	45.449	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.900%
35) Dibromofluoromethane	8.172	113	231258	46.714	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.420%
50) Toluene-d8	10.572	98	876300	48.029	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.060%
62) 4-Bromofluorobenzene	12.854	95	317697	47.606	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.220%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.143	85	7336	0.880	ug/l	92
9) 1,1,2-Trichlorotrifluo...	4.384	101	72542	13.065	ug/l	97
12) 1,1-Dichloroethene	4.360	96	28095	5.490	ug/l	98
24) 1,1-Dichloroethane	6.578	63	6862	0.661	ug/l #	80
27) cis-1,2-Dichloroethene	7.495	96	10617	1.696	ug/l	98
30) Chloroform	7.972	83	15233	1.367	ug/l	97
38) Carbon Tetrachloride	8.372	117	16956	1.892	ug/l	95
44) Trichloroethene	9.360	130	1339195	229.606	ug/l	99
49) 1,4-Dioxane	9.713	88	1827	18.219	ug/l #	65
52) Toluene	10.636	92	3545	0.255	ug/l	88
55) 1,1,2-Trichloroethane	11.019	97	3782	0.700	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076964.D
Acq On : 13 Mar 2023 16:08
Operator : JC\MD
Sample : 01880-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE115D1-20230309

Quant Time: Mar 14 07:03:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

