

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076965.D
 Acq On : 13 Mar 2023 16:32
 Operator : JC\MD
 Sample : 01880-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT14911-20230309

Quant Time: Mar 14 07:04:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.237	168	398515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	629582	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	550410	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	228205	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	268538	45.156	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.320%
35) Dibromofluoromethane	8.172	113	201227	47.957	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.920%
50) Toluene-d8	10.572	98	745677	48.220	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.440%
62) 4-Bromofluorobenzene	12.854	95	267981	47.378	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.760%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.354	130	4871	0.985	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076965.D
Acq On : 13 Mar 2023 16:32
Operator : JC\MD
Sample : 01880-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT14911-20230309

Quant Time: Mar 14 07:04:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

