

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076966.D
 Acq On : 13 Mar 2023 16:56
 Operator : JC\MD
 Sample : 01880-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT149S1-20230309

Quant Time: Mar 14 07:04:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	380878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	613888	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	522047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	207724	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	269358	47.391	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.780%
35) Dibromofluoromethane	8.172	113	190974	46.677	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.360%
50) Toluene-d8	10.571	98	716466	47.515	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.040%
62) 4-Bromofluorobenzene	12.854	95	252292	45.744	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.480%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	5.813	73	156177	10.433	ug/l	97
42) 1,2-Dichloroethane	8.672	62	3570	0.491	ug/l	75
44) Trichloroethene	9.360	130	20081	4.166	ug/l	96
64) Tetrachloroethene	11.101	164	3212	0.760	ug/l	94
91) 1,2-Dichlorobenzene	14.107	146	1840	0.243	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076966.D
Acq On : 13 Mar 2023 16:56
Operator : JC\MD
Sample : 01880-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT149S1-20230309

Quant Time: Mar 14 07:04:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

