

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076967.D
 Acq On : 13 Mar 2023 17:21
 Operator : JC\MD
 Sample : 01880-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT158S1-20230309

Quant Time: Mar 14 07:04:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	375675	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	600108	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	529343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	216601	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	261084	46.571	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.140%
35) Dibromofluoromethane	8.178	113	186019	46.510	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.020%
50) Toluene-d8	10.572	98	713151	48.381	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.760%
62) 4-Bromofluorobenzene	12.854	95	253214	46.966	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.940%
Target Compounds						
					Qvalue	
24) 1,1-Dichloroethane	6.584	63	11627	1.381	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076967.D
Acq On : 13 Mar 2023 17:21
Operator : JC\MD
Sample : 01880-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT158S1-20230309

Quant Time: Mar 14 07:04:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

