

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076968.D
 Acq On : 13 Mar 2023 17:45
 Operator : JC\MD
 Sample : 01880-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT15811-20230309

Quant Time: Mar 14 07:05:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	382126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	614962	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	539368	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	232188	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	265826	46.617	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.240%
35) Dibromofluoromethane	8.178	113	193710	47.263	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.520%
50) Toluene-d8	10.571	98	717479	47.499	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.000%
62) 4-Bromofluorobenzene	12.854	95	264153	47.811	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.620%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.360	130	1460	0.302	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076968.D
Acq On : 13 Mar 2023 17:45
Operator : JC\MD
Sample : 01880-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT15811-20230309

Quant Time: Mar 14 07:05:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

