

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076969.D
 Acq On : 13 Mar 2023 18:09
 Operator : JC\MD
 Sample : 01880-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE105D1-20230309

Quant Time: Mar 14 07:05:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	341124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	544821	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	484085	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	211545	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	244896	48.108	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.220%
35) Dibromofluoromethane	8.178	113	170828	47.046	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.100%
50) Toluene-d8	10.572	98	643635	48.096	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.200%
62) 4-Bromofluorobenzene	12.854	95	240136	49.060	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.120%
Target Compounds						
2) Dichlorodifluoromethane	2.137	85	2289	0.373	ug/l #	49
9) 1,1,2-Trichlorotrifluo...	4.390	101	20415	4.990	ug/l	98
12) 1,1-Dichloroethene	4.360	96	2774	0.736	ug/l	96
27) cis-1,2-Dichloroethene	7.495	96	3552	0.770	ug/l	97
32) 1,1,1-Trichloroethane	8.183	97	4030	0.523	ug/l #	51
44) Trichloroethene	9.360	130	195474	45.693	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076969.D
Acq On : 13 Mar 2023 18:09
Operator : JC\MD
Sample : 01880-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE105D1-20230309

Quant Time: Mar 14 07:05:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

