

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076971.D
 Acq On : 13 Mar 2023 18:57
 Operator : JC\MD
 Sample : 01880-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D3-20230310

Quant Time: Mar 14 07:06:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	426453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	669973	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	583767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	247383	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	302784	47.579	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.160%		
35) Dibromofluoromethane	8.172	113	215902	48.353	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.700%		
50) Toluene-d8	10.572	98	785083	47.707	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.420%		
62) 4-Bromofluorobenzene	12.854	95	287686	47.795	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.600%		
Target Compounds					Qvalue	
44) Trichloroethene	9.354	130	22153	4.211	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076971.D
Acq On : 13 Mar 2023 18:57
Operator : JC\MD
Sample : 01880-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D3-20230310

Quant Time: Mar 14 07:06:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

