

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076972.D
 Acq On : 13 Mar 2023 19:21
 Operator : JC\MD
 Sample : 01880-10 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE105D2-20230309

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/14/2023
 Supervised By : Mahesh Dadoda 03/14/2023

Quant Time: Mar 14 07:06:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	339817	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	543677	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	490377	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	212653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	248022	48.910	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.820%		
35) Dibromofluoromethane	8.172	113	176623	48.745	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.480%		
50) Toluene-d8	10.572	98	648907	48.592	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.180%		
62) 4-Bromofluorobenzene	12.854	95	238648	48.858	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.720%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.366	101	2438m	0.598	ug/l	Qvalue
44) Trichloroethene	9.360	130	282512	66.177	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076972.D
Acq On : 13 Mar 2023 19:21
Operator : JC\MD
Sample : 01880-10 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE105D2-20230309

Quant Time: Mar 14 07:06:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 03/14/2023
Supervised By :Mahesh Dadoda 03/14/2023

