

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
 Data File : VN076973.D
 Acq On : 13 Mar 2023 19:46
 Operator : JC\MD
 Sample : 01880-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D2-20230310

Quant Time: Mar 14 07:07:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 10 13:47:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	316733	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	490725	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	443966	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	195804	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	230849	48.841	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.680%
35) Dibromofluoromethane	8.172	113	156741	47.925	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.860%
50) Toluene-d8	10.572	98	587158	48.713	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.420%
62) 4-Bromofluorobenzene	12.854	95	216500	49.107	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.220%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.149	85	1735	0.304	ug/l	76
9) 1,1,2-Trichlorotrifluo...	4.390	101	92980	24.476	ug/l	95
12) 1,1-Dichloroethene	4.360	96	6819	1.948	ug/l #	74
27) cis-1,2-Dichloroethene	7.501	96	10400	2.428	ug/l	96
30) Chloroform	7.978	83	3759	0.493	ug/l #	83
38) Carbon Tetrachloride	8.366	117	4967	0.839	ug/l #	82
44) Trichloroethene	9.360	130	1978005	513.337	ug/l	100
49) 1,4-Dioxane	9.707	88	1406	21.223	ug/l #	96
55) 1,1,2-Trichloroethane	11.024	97	1343	0.376	ug/l #	67
64) Tetrachloroethene	11.107	164	22097	6.150	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031323\
Data File : VN076973.D
Acq On : 13 Mar 2023 19:46
Operator : JC\MD
Sample : 01880-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D2-20230310

Quant Time: Mar 14 07:07:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031023W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 10 13:47:39 2023
Response via : Initial Calibration

