

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031325\
 Data File : VN085972.D
 Acq On : 13 Mar 2025 11:38
 Operator : JC\MD
 Sample : VN0313WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0313WBL01

Quant Time: Mar 14 00:27:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	143726	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	265668	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	228391	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	91345	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	108246	58.092	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.180%
35) Dibromofluoromethane	8.165	113	89964	51.751	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.500%
50) Toluene-d8	10.565	98	291558	46.097	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.200%
62) 4-Bromofluorobenzene	12.847	95	91420	43.870	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031325\
Data File : VN085972.D
Acq On : 13 Mar 2025 11:38
Operator : JC\MD
Sample : VN0313WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0313WBL01

Quant Time: Mar 14 00:27:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

