

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031325\
 Data File : VN085976.D
 Acq On : 13 Mar 2025 13:24
 Operator : JC\MD
 Sample : PB167104ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167104ZHE#01

Quant Time: Mar 14 00:28:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	137719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	254569	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	253553	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	96515	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	111282	62.326	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 124.660%		
35) Dibromofluoromethane	8.165	113	90827	54.525	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 109.060%		
50) Toluene-d8	10.565	98	336919	55.592	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 111.180%		
62) 4-Bromofluorobenzene	12.847	95	104527	52.346	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 104.700%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	40067	70.920	ug/l	97
20) Methylene Chloride	5.283	84	5601	3.083	ug/l #	81
43) Isopropyl Acetate	8.688	43	53147	15.229	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031325\
Data File : VN085976.D
Acq On : 13 Mar 2025 13:24
Operator : JC\MD
Sample : PB167104ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167104ZHE#01

Quant Time: Mar 14 00:28:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

