

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031325\
 Data File : VN085982.D
 Acq On : 13 Mar 2025 16:03
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 14 00:29:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	48#	0.00
2 T	Dichlorodifluoromethane	0.680	0.725	-6.6	52	0.00
3 P	Chloromethane	0.682	0.578	15.2	45#	0.00
4 C	Vinyl Chloride	0.723	0.602	16.7#	43#	0.00
5 T	Bromomethane	0.462	0.378	18.2	43#	0.00
6 T	Chloroethane	0.479	0.362	24.4	42#	0.00
7 T	Trichlorofluoromethane	1.046	1.058	-1.1	53	0.00
8 T	Diethyl Ether	0.339	0.334	1.5	49#	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.605	0.595	1.7	51	0.01
10 T	Methyl Iodide	0.745	0.764	-2.6	51	0.00
11 T	Tert butyl alcohol	0.080	0.072	10.0	44#	0.01
12 CM	1,1-Dichloroethene	0.548	0.545	0.5#	50#	0.00
13 T	Acrolein	0.118	0.119	-0.8	45#	0.00
14 T	Allyl chloride	0.720	0.686	4.7	48#	0.00
15 T	Acrylonitrile	0.259	0.255	1.5	50	0.00
16 T	Acetone	0.205	0.221	-7.8	58	0.00
17 T	Carbon Disulfide	1.625	1.612	0.8	53	0.00
18 T	Methyl Acetate	0.705	0.552	21.7	43#	0.00
19 T	Methyl tert-butyl Ether	1.629	1.806	-10.9	52	0.00
20 T	Methylene Chloride	0.660	0.633	4.1	51	0.00
21 T	trans-1,2-Dichloroethene	0.590	0.591	-0.2	52	0.00
22 T	Diisopropyl ether	1.627	1.617	0.6	47#	0.00
23 T	Vinyl Acetate	1.114	1.143	-2.6	49#	0.00
24 P	1,1-Dichloroethane	1.106	1.068	3.4	50#	0.00
25 T	2-Butanone	0.302	0.298	1.3	49#	0.00
26 T	2,2-Dichloropropane	0.983	1.035	-5.3	53	0.00
27 T	cis-1,2-Dichloroethene	0.685	0.684	0.1	50#	0.00
28 T	Bromochloromethane	0.454	0.470	-3.5	51	0.00
29 T	Tetrahydrofuran	0.195	0.194	0.5	48#	0.00
30 C	Chloroform	1.156	1.173	-1.5#	53	0.00
31 T	Cyclohexane	0.926	0.903	2.5	51	0.00
32 T	1,1,1-Trichloroethane	1.028	1.094	-6.4	54	0.00
33 S	1,2-Dichloroethane-d4	0.648	0.627	3.2	49#	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	47#	0.00
35 S	Dibromofluoromethane	0.327	0.314	4.0	46#	0.00
36 T	1,1-Dichloropropene	0.463	0.505	-9.1	52	0.00
37 T	Ethyl Acetate	0.381	0.373	2.1	47#	0.00
38 T	Carbon Tetrachloride	0.550	0.637	-15.8	57	0.00
39 T	Methylcyclohexane	0.455	0.517	-13.6	49#	0.00
40 TM	Benzene	1.484	1.585	-6.8	52	0.00
41 T	Methacrylonitrile	0.204	0.196	3.9	43#	0.00
42 TM	1,2-Dichloroethane	0.483	0.538	-11.4	55	0.00
43 T	Isopropyl Acetate	0.807	0.706	12.5	52	0.00
44 TM	Trichloroethene	0.364	0.361	0.8	50	0.00
45 C	1,2-Dichloropropane	0.358	0.364	-1.7#	49#	0.00
46 T	Dibromomethane	0.253	0.269	-6.3	52	0.00
47 T	Bromodichloromethane	0.541	0.588	-8.7	53	0.00
48 T	Methyl methacrylate	0.283	0.299	-5.7	47#	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031325\
 Data File : VN085982.D
 Acq On : 13 Mar 2025 16:03
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 14 00:29:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.006	-20.0	50	0.00
50 S	Toluene-d8	1.190	1.165	2.1	46#	0.00
51 T	4-Methyl-2-Pentanone	0.367	0.392	-6.8	49#	0.00
52 CM	Toluene	0.870	1.008	-15.9#	53	0.00
53 T	t-1,3-Dichloropropene	0.500	0.582	-16.4	52	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.618	-12.8	51	0.00
55 T	1,1,2-Trichloroethane	0.346	0.370	-6.9	53	0.00
56 T	Ethyl methacrylate	0.434	0.546	-25.8#	52	0.00
57 T	1,3-Dichloropropane	0.577	0.622	-7.8	52	0.00
58 T	2-Chloroethyl Vinyl ether	0.195	0.212	-8.7	57	0.00
59 T	2-Hexanone	0.253	0.289	-14.2	50#	0.00
60 T	Dibromochloromethane	0.403	0.462	-14.6	54	0.00
61 T	1,2-Dibromoethane	0.327	0.365	-11.6	53	0.00
62 S	4-Bromofluorobenzene	0.392	0.433	-10.5	49#	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	50	0.00
64 T	Tetrachloroethene	0.387	0.372	3.9	52	0.00
65 PM	Chlorobenzene	1.145	1.178	-2.9	53	0.00
66 T	1,1,1,2-Tetrachloroethane	0.412	0.435	-5.6	55	0.00
67 C	Ethyl Benzene	1.794	2.049	-14.2#	54	0.00
68 T	m/p-Xylenes	0.682	0.814	-19.4	55	0.00
69 T	o-Xylene	0.648	0.762	-17.6	54	0.00
70 T	Styrene	1.059	1.335	-26.1#	56	0.00
71 P	Bromoform	0.306	0.345	-12.7	57	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	55	0.00
73 T	Isopropylbenzene	3.425	3.572	-4.3	55	0.00
74 T	N-amyl acetate	1.208	1.149	4.9	51	0.00
75 P	1,1,2,2-Tetrachloroethane	1.179	1.042	11.6	55	0.00
76 T	1,2,3-Trichloropropane	1.111	0.975	12.2	57	0.00
77 T	Bromobenzene	0.925	0.916	1.0	56	0.00
78 T	n-propylbenzene	3.923	4.256	-8.5	55	0.00
79 T	2-Chlorotoluene	2.621	2.656	-1.3	56	0.00
80 T	1,3,5-Trimethylbenzene	2.790	3.143	-12.7	57	0.00
81 T	trans-1,4-Dichloro-2-butene	0.381	0.414	-8.7	58	0.00
82 T	4-Chlorotoluene	2.598	2.764	-6.4	58	0.00
83 T	tert-Butylbenzene	2.346	2.552	-8.8	55	0.00
84 T	1,2,4-Trimethylbenzene	2.761	3.252	-17.8	58	0.00
85 T	sec-Butylbenzene	3.315	3.598	-8.5	54	0.00
86 T	p-Isopropyltoluene	2.665	3.089	-15.9	56	0.00
87 T	1,3-Dichlorobenzene	1.734	1.733	0.1	59	0.00
88 T	1,4-Dichlorobenzene	1.777	1.706	4.0	57	0.00
89 T	n-Butylbenzene	2.354	2.522	-7.1	55	0.00
90 T	Hexachloroethane	0.628	0.609	3.0	57	0.00
91 T	1,2-Dichlorobenzene	1.667	1.651	1.0	58	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.216	0.202	6.5	58	0.00
93 T	1,2,4-Trichlorobenzene	0.794	0.810	-2.0	56	0.00
94 T	Hexachlorobutadiene	0.483	0.439	9.1	57	0.00
95 T	Naphthalene	2.181	2.243	-2.8	53	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031325\
Data File : VN085982.D
Acq On : 13 Mar 2025 16:03
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 14 00:29:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.798	0.801	-0.4	56	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6