

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031419\
 Data File : VN054259.D
 Acq On : 14 Mar 2019 14:01
 Operator : JC/SP
 Sample : K1899-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Mar 15 07:26:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1305655	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2140251	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2241019	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	976143	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	696234	50.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.42%	
35) Dibromofluoromethane	7.59	113	666638	48.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.58%	
50) Toluene-d8	10.09	98	2684603	52.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.32%	
62) 4-Bromofluorobenzene	12.40	95	1014702	56.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.78%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN031419\
Data File : VN054259.D
Acq On : 14 Mar 2019 14:01
Operator : JC/SP
Sample : K1899-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Mar 15 07:26:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

