

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031419\
 Data File : VN054273.D
 Acq On : 14 Mar 2019 20:21
 Operator : JC/SP
 Sample : K1887-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Mar 15 08:04:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1396407	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2261734	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2314054	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	900474	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	715208	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
35) Dibromofluoromethane	7.59	113	701002	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
50) Toluene-d8	10.09	98	2772182	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
62) 4-Bromofluorobenzene	12.40	95	1017445	53.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.96%	
Target Compounds						
95) Naphthalene	15.13	128	47748	4.595	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN031419\
Data File : VN054273.D
Acq On : 14 Mar 2019 20:21
Operator : JC/SP
Sample : K1887-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Mar 15 08:04:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

