

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031419\
 Data File : VN054280.D
 Acq On : 15 Mar 2019 00:27
 Operator : JC/SP
 Sample : VN0314WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0314WBL02

Quant Time: Mar 15 08:16:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1340319	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2204938	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2317232	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	941761	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	707645	49.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.42%	
35) Dibromofluoromethane	7.59	113	678297	47.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.40%	
50) Toluene-d8	10.09	98	2737480	52.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.24%	
62) 4-Bromofluorobenzene	12.40	95	1035651	56.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN031419\
Data File : VN054280.D
Acq On : 15 Mar 2019 00:27
Operator : JC/SP
Sample : VN0314WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0314WBL02

Quant Time: Mar 15 08:16:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

