

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031419\
 Data File : VN054284.D
 Acq On : 15 Mar 2019 2:16
 Operator : JC/SP
 Sample : PB117849TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117849TB

Quant Time: Mar 15 08:24:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1352326	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2250600	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2361744	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	982599	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	717384	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
35) Dibromofluoromethane	7.59	113	694923	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
50) Toluene-d8	10.09	98	2814298	52.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.00%	
62) 4-Bromofluorobenzene	12.40	95	1054053	56.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.38%	
Target Compounds						
20) Methylene Chloride	4.55	84	19860	1.471	ug/l #	88
95) Naphthalene	15.14	128	15265	3.799	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN031419\
Data File : VN054284.D
Acq On : 15 Mar 2019 2:16
Operator : JC/SP
Sample : PB117849TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117849TB

Quant Time: Mar 15 08:24:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

