

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN031419\
 Data File : VN054302.D
 Acq On : 15 Mar 2019 10:29
 Operator : JC/SP
 Sample : K1902-12 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30240-41-42-43-44

Quant Time: Mar 15 18:47:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 13 03:33:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1431965	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2364693	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	2479502	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	1023513	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	727786	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
35) Dibromofluoromethane	7.59	113	713406	46.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.56%	
50) Toluene-d8	10.09	98	2921848	51.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.74%	
62) 4-Bromofluorobenzene	12.40	95	1108465	56.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.48%	
Target Compounds						
68) m/p-Xylenes	11.62	106	7079	0.201	ug/l	89
84) 1,2,4-Trimethylbenzene	13.04	105	26592	0.398	ug/l	94
95) Naphthalene	15.13	128	12140	3.721	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN031419\
Data File : VN054302.D
Acq On : 15 Mar 2019 10:29
Operator : JC/SP
Sample : K1902-12 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30240-41-42-43-44

Quant Time: Mar 15 18:47:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031219W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 13 03:33:31 2019
Response via : Initial Calibration

