

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
 Data File : VN071262.D
 Acq On : 14 Mar 2022 13:38
 Operator : JC\MD
 Sample : PB143286TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143286TB

Quant Time: Mar 15 02:51:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	777091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1279020	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1184686	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	437723	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	492437	46.237	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.480%
35) Dibromofluoromethane	8.016	113	380020	47.084	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.160%
50) Toluene-d8	10.439	98	1567653	48.241	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.480%
62) 4-Bromofluorobenzene	12.727	95	575842	52.330	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.660%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	36011	6.949	ug/l	92
18) Methyl Acetate	4.857	43	26533	2.359	ug/l	92
20) Methylene Chloride	5.092	84	15211	1.976	ug/l #	87
43) Isopropyl Acetate	8.557	43	119612	4.494	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN031422\
Data File : VN071262.D
Acq On : 14 Mar 2022 13:38
Operator : JC\MD
Sample : PB143286TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143286TB

Quant Time: Mar 15 02:51:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

